

SmartLinx Axon Regulatory Notice

Introduction

This Notice contains regulatory information for the SmartLinx Axon product and is therefore an extension of the *SmartLinx Axon Instructions for Use*. It details the compliance statements that the product requires for its certification and approval. Capsule Technologie is therefore committed to delivering products compliant with standards, laws and regulations.

Certification Marks



Australia Regulatory Compliance Mark (RCM). The SmartLinx Axon complies with the Radiocommunication Act 1992.



Compliance with European directives



Separate collection for electrical and electronic waste (WEEE directive)



Gulf Mark (G-Mark) for compliance with Gulf Technical Regulation for Low Voltage Electrical Equipment and Appliances.



NRTL safety mark



Compliance to Mexican standards



Compliance for products being placed in Great Britain.

Regulatory compliance and approvals

Federal Communications Commission

SmartLinx Axon complies with Part 15 of the FCC Rules. Operation is subject to two conditions:

- This device may not cause harmful interference.
- This device must accept any interference received, including interference that may cause undesired operation.

Note: This equipment has been tested and found to comply with the limits for a Class A digital device, pursuant to Part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference when the equipment is operated in a commercial environment. This equipment generates, uses, and can radiate radio frequency energy and, if not installed and used in accordance with the instruction manual, may cause harmful interference to radio communications. Operation of this equipment in a residential area is likely to cause harmful interference in which case the user will be required to correct the interference at his own expense.

Note: For operation within 5.180 ~ 5.250GHz / 5.500 ~5.700 GHz frequency range, SmartLinX Axon is restricted to indoor environment. The band from 5600-5650 MHz will be disabled by the software during the manufacturing and cannot be changed by the end user. This device meets all the other requirements specified in Part 15E, Section 15.407 of the FCC Rules.

Caution: Capsule is not responsible for any radio or television interference caused by unauthorized changes or modifications to this equipment. Unauthorized changes or modifications could void the user's authority to operate the equipment.

Caution: Any changes or modifications not expressly approved by the party responsible for compliance could void the user's authority to operate this equipment. This transmitter must not be co-located or operating in conjunction with any other antenna or transmitter.

Radiation Exposure Statement

This equipment complies with FCC radiation exposure limits set forth for an uncontrolled environment. This equipment should be installed and operated with minimum distance 20 cm between the radiator and your body.

Note: The country code selection is for non-US model only and is not available on all US models. Per FCC regulation, all Wi-Fi products marketed in the USA must fixed to US-operation channels only.

Great Britain

For a copy of the full UKCA Declaration of Conformity, contact your Capsule Technologie representative.

RoHS compliance

Capsule complies with the restriction of the use of certain hazardous substances in Electrical and Electronic Equipment Regulations 2012 as explained in *RoHS and WEEE compliance* below.

Industry Canada statement

This device complies with RSS-210 of the Industry Canada Rules. Operation is subject to the following two conditions: (1) This device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.

- Caution:**
- i. the device for operation in the band 5180-5250 MHz is only for indoor use to reduce the potential for harmful interference to co-channel mobile satellite systems;
 - ii. the maximum antenna gain permitted for devices in the bands 5260-5320 MHz and 5500-5700 MHz shall comply with the e.i.r.p. limit; and
 - iii. the maximum antenna gain permitted for devices in the band 5500-5700 MHz shall comply with the e.i.r.p. limits specified for point-to-point and non point-to-point operation as appropriate.
 - iv. users should also be advised that high-power radars are allocated as primary users (i.e. priority users) of the bands 5260-5320 MHz and 5500-5700 MHz and that these radars could cause interference and/or damage to LE-LAN devices.

Radiation Exposure Statement

This equipment complies with IC radiation exposure limits set forth for an uncontrolled environment. This equipment should be installed and operated with minimum distance 20cm between the radiator and your body.

Mexico certification

The Smartlinx Axon complies with the official Mexican standards – NOM (Normas Oficiales Mexicanas) :

- NOM-019-SCFI-1998: Data Processing Equipment – Safety requirements
- NOM-208-SCFI-2016: Products. Radiocommunication systems that employ the dispersed spectrum technique –Radiocommunication equipment – Frequency hopping radiocommunication and digital modulation equipment to operate in the 902 MHz-928 MHz, 2400 MHz-2483.5 MHz and 5725 MHz-5850 MHz-bands – Specifications and test methods
- IFT-008-2015: Radiocommunication systems that use spread-spectrum services - Radiocommunication equipments with frequency hopping and digital modulation that operate in the bands 902MHz-928MHz, 2400MHz-2483, 5MHz and 5725MHz-5850MHZ - Specifications, limits

The SmartLinx Axon is issued an IFETEL ID number depending on its type:

- Axon 110 : RCPCASL21-1025
- Axon 410 : RCPCASL21-1024
- Axon 810 : RCPCASL21-1023

REACH compliance

Pursuant to REACH regulation 1907/2006, Article 33, and the introduction of Lead in the list of Substances of Very High Concern (SVHC) by the European Chemical Agency (ECHA), Capsule Technologie has identified a component in the SmartLinx Axon containing lead above a 0.1% weight/weight.

Lead is detected in a component of the SmartLinx Axon 110 (SL-AXON110-HW) internal power supply. This component is internal and consequently does not expose users to the substance. Moreover, this component is not intended to release its substance under normal or reasonably foreseeable conditions of use.

SmartLinx Axon 410 (SL-AXON410-HW) and SmartLinx Axon 810 (SL-AXON810-HW) do not contain SVHC.

All of the products listed above comply with Directive RoHS 2011/65/EU, as explained in the next section (*RoHS and WEEE compliance*).

RoHS and WEEE compliance

The SmartLinx Axon does NOT contain any of the following substances (in concentrations exceeding legal threshold limits):

- Lead
- Mercury
- Cadmium
- Hexavalent Chromium
- Polybrominated Biphenyls (PBB)
- Polybrominated Diphenyl Ethers (PBDE)
- Bis (2-Ethylhexyl) phthalate (DEHP)
- Benzyl butyl phthalate (BBP)
- Dibutyl phthalate (DBP)
- Diisobutyl phthalate (DIBP)

SmartLinx Axons and cables should be collected separately and not disposed of with household waste. Refer to the section on *Recycling and the Environment* for more details.

Note: For more information about the Japan RoHS compliance, refer to the website at <https://www.capsuletech.com/notices>.

Recycling and the environment

Improper disposal of IT equipment can have a negative impact on health and the environment. We recommend that you dispose of SmartLinx Axon, DIM, and serial cables at an appropriate facility to enable recovery and recycling. You can also recycle packaging and guides according to your local recycling regulations.

In the European Union, for assistance with the recycling of Capsule products, visit our customer site:

<https://customers.capsuletech.com/environment>

Safety and Regulatory Compliance table

Field	Standard or regulation
Medical device safety	EN 60601-1 IEC 60601-1 3rd edition with national deviations for USA and Canada
Medical device usability	IEC 60601-1-6 IEC 62366
Medical device Software – Software Lifecycle Processes	IEC 62304
EMC/EMI	FCC 47 CFR Part 15 sub-part B ICES-003 A / NMB-003 A EN 60601-1-2 CISPR 24

Field	Standard or regulation
	CISPR 32
Radio	EN 301 489-1 EN 301 489-17 EN 301 893 EN 300 328 EN 62311 RSS-210 FCC 47 CFR 15 C FCC 47 CFR 15 E
OET65	ANSI/IEEE C95.1
Environment/Packaging	EU Directive 94/62/EC
Environment	REACH 1907/2006
RoHS	EU Directive 2011/65/EU
WEEE	EU Directive 2012/19/EU

Japan Ministry of Internal Affairs and Communications (MIC)

The SmartLinx Axon is certified to the Technical Regulation Conformity Certification under the Japanese Radio Law.

KSA Communications and Information Technology Commission (CITC)

The SmartLinx Axon complies with the Wireless Local Area Networks Regulation for the Kingdom of Saudi Arabia.

Info-communications Media Development Authority (IMDA) Singapore

The SmartLinx Axon is an IMDA registered product.

Independent Communications Authority of South Africa (ICASA)

The SmartLinx Axon complies with the ICASA Act, 2000.

United Arab Emirates (UAE)

For the Telecommunication Regulatory Authority (TRA) approval:

- Capsule Tech SAS is a TRA registered dealer (DA38012/15).
- The SmartLinx Axon are TRA registered equipment:
 - SmartLinx Axon 110 (ER55369/17)
 - SmartLinx Axon 410 (ER55368/17)
 - SmartLinx Axon 810 (ER55367/17)

Legal Notice

© 2019, 2021 Capsule Technologies, Inc. And/or its subsidiaries. All rights reserved.

This product and related documentation are protected by copyright and distributed under licensing restricting their use, copying, distribution, and decompilation. No part of this product or related documentation may be reproduced in any form or by any means without prior written authorization of Capsule Technologies, Inc. And/or its subsidiaries.

This technical data may be subject to U.S. And international export, re-export, or transfer ("export") laws. Diversion contrary to U.S. And international law is strictly prohibited.

The Capsule logo is a registered trademark of Capsule Technologie, SAS in the US, EU and other countries. All other trademarks, service marks, registered trademarks and registered service marks are the property of their respective owners.

The hardware and software of Capsule Technologies, Inc. And/or its subsidiaries is provided subject to all third-party licenses and limitations in the "third party notices" electronic file included as part of the software or available upon request from Capsule Technologies, Inc. And/or its subsidiaries.

DHF18688

DCN 2021-039